

S. No.	Item	Specifications	Vendors Participated			
			ARTECH	Medinostic	Shazup	Medline
1	CPR Mannequin (Child)	Manikin must be compatible with latest AHA guidelines and should be upgradable with electronic feedback option. It must be enabled with Bluetooth connectivity for feedback option. Any mobile/tablet (Android/iOS) can be used as feedback device along-with following features: • Real-time feedback on compressions and ventilations • Post-training debriefs, including overall scores and tips for improvement • Details on compression release, depth and rate, ventilation volume, and number of compressions, ventilations and cycles.	Not Quoted as per BOQs Disqualified <i>(I)Real-time feedback on compression and ventilations are absent (ref. pre bid document) (II)It does not have post-training debriefs,including overall scores and tips for improvement as given in specification(ref.pre bid doc)(III) It doesnot offer details on copression release, depth and rate, ventilation volume, and number of compression, ventilations and cycles (ref. to pre. bid doc.)</i>	Quoted as per BOQs Qualified		
2	BVM (Child)	Mentioned in BOQs	Quoted as per BOQs Qualified	Quoted as per BOQs Qualified	-	-
3	BVM (Infant)	Mentioned in BOQs	Quoted as per BOQs Qualified	Quoted as per BOQs Qualified	-	-
4	Set of Oropharyngeal Airways (OPA)	Mentioned in BOQs	-	-	-	-
5	CPR Mouth Peace	Mentioned in BOQs	-	Quoted as per BOQs Qualified	-	-
6	Spine Board + Head Immobilizer	It should comply for following features: • Floatable • X Ray Translucent • built in side handles • size 6 feet x 1.5 feet • fluid prevent contamination • load capacity over 180 Kg • Protection straps • Adjustable Head immobilizer for all sizes • Latex Free • X Ray Compatible • Must be CE / FDA / EN Certified	-	Quoted as per BOQs Qualified	-	-

Dr. Mansoor Chani

Programme Coordinator

MSC Nursing

University of Health Sciences Lahore

7	Spider Webbing	Mentioned in BOOs	-	-	-	-	-	-	-
8	Orthopedic Stretcher	Mentioned in BOOs	-	-	-	-	-	-	-


 Dr. Mansoor Chani
 Program Coordinator
 MSc Nursing
 University of Health Sciences Lahore

9	Vacuum Splints	Mentioned in BOQs	-	-	-	-	-
10	Pack of Frac Straps	Mentioned in BOQs	-	-	-	-	-
11	Motorcycle Helmet	Mentioned in BOQs	-	-	-	-	-
12	Triangular bandages	Mentioned in BOQs	-	-	-	-	-
13	Pressure Bandages	Mentioned in BOQs	-	-	-	-	-
14	Oxygen Cylinder	Mentioned in BOQs	-	-	-	-	-
15	Portable Oxygen Cylinder	Mentioned in BOQs	-	-	-	-	-
16	Treache Intubation Module	Mentioned in BOQs	-	-	-	-	-
17	Nasal Cannula	Mentioned in BOQs	-	-	-	-	-
18	Torch	Mentioned in BOQs	-	-	-	-	-
19	Defibrillator Machine	● Compact 4-in-1 unit Manual Defibrillation Automatic External Defibrillation Non-Invasive Pacing Effective CPR solution ● Technology: Biphasic truncated Exponential (BTE) waveform, auto-compensation according to patient impedance ● 7" TFT Color display with 3 waveforms assures easy viewing ● Escalating dose of 1-360J to maximize defibrillation success ● Display of HR, ECG Through paddles and leads I, II, III P patient cable ● Synchronized output with ECG ● Built in recorder for printing of full summary on 50mm paper ● 5 segment LED Charging indicator ● 7.5 Seconds for whole defibrillation operation - Power on in 2 seconds - Charge to shock in 3 seconds - ECG recovery in only 2.5 seconds ● Alarm for high and low heart rate, low battery warning ● Built in Rechargeable battery with charger ● Auto tester /Self check ● Intuitive contact impedance indicator ● 120 Seconds ECG Waveform Review ● Powerful data storage for up to 100 patient profiles, 1,000 events per patient, 24 hours of ECG waveforms, and 72h trend analysis	-	Not Quoted as per BOQs Disqualified (I) Non Invasive pacing absent (ref. pre bid doc) (II) Effective CPR solution absent (ref. pre bid doc) (III) It does not reach Escalating dose of 1-360J to maximize defibrillation success (IV) It takes more than 7.5 seconds for whole defibrillation operation which is against specifications (ref. to pre bid doc) (V) 120 Seconds ECG Waveform review absent (ref. to pre bid doc) (VI) It does not have powerful data storage for up to 100 patient profiles, 1,000 events per patient 24 hours of ECG waveforms and 72h trend analysis as per given specifications (ref. to pre bid doc.)	Quoted as per BOQs	Quoted as per BOQs	Not Quoted as per BOQs

Dr. Mansoor Ghani
Programme Coordinator
MSC Nursing
University of Health Sciences Lahore

20	Precision Blood Glucose Kit	Mentioned in BOQs	-	-	Quoted as per BOQs <i>Qualified</i>	-	-
21	Suction Unit	Mentioned in BOQs	-	-	Quoted as per BOQs <i>Qualified</i>	-	-
22	Yankaur Catheter	Mentioned in BOQs	-	-	-	-	-
23	Non-Rebreather Mask	Mentioned in BOQs	-	-	-	-	-
24	NP Airway	Mentioned in BOQs	-	-	-	-	-
25	Laryngoscope	Mentioned in BOQs	-	-	Quoted as per BOQs <i>Qualified</i>	-	-
26	LMA 3,4,5 sizes	Mentioned in BOQs	-	-	-	-	-
27	I Gel 3,4,5	Mentioned in BOQs	-	-	-	-	-
28	Catheter mount	Mentioned in BOQs	-	-	-	-	-
29	CO2 detector	Mentioned in BOQs	-	-	-	-	-
30	Vocal cord model	Mentioned in BOQs	-	-	-	-	-


 Dr. Mansoor Ghani
 Programme Coordinator
 MSC Nursing
 University of Health Sciences Lahore


 Dr. Mansoor Ghani
 Programme Coordinator
 MSC Nursing
 University of Health Sciences Lahore