## University of Health Sciences

Khayaban – e – Jamia Punjab Lahore – 54600, Pakistan Fax: +92 42 9230394 www.uhs.edu.pk



## **Department of Microbiology**

Phone: 111 333 366 Ext. 355 e-mail: microbiology@uhs.edu.pk

Dated: 25-07-2019

Project Coordinator SARF-KSK

# Subject: <u>Strengthening of Academic and Research Facilities at Kala Shah Kaku:</u> <u>Corrections regarding technical evaluation report</u>

- 1. Sr No. 6 & 7 i.e. Hot air oven and pH meter are not quoted by "ScienTec" in the technical bid provided to the department.
- 2. Pakistan Hospital has only quoted three adjustable Micropippetes i.e.100-1000 $\mu$ l, 10-100  $\mu$ l & 1-10  $\mu$ l, however the vendor has not quoted the fourth i.e. 0.5-10  $\mu$ l in the technical bid provided, those quoted have been added to the technical evaluation report.

Dr. Sidrah Saleem

Head Department of Microbiology

#### University of Health Sciences

Khayaban-e-Jamia Punjab, 54600, Lahore, Pakistan Ph: +92 42 99231304-9, Fax: +92 42 99230870 www.us.edu.pk



## Department of Human Genetics & Molecular Biology

Ext: 353, 320

Ref. No: UHS/H4MB/H/982

Dated: 24th July, 2019

The Project Coordinator UHS/HGMB/7/19/ Project SARF-KSK University of Health Science, Lahore.

## Subject: Grievances regarding Technical evaluation report for KSK-SARF tender (Lab. Equipments)

With reference to the notification dated 23-07-2019, please find below the responses to grievances received for technical evaluation report, KSK-SARF tender, concerning items from the **Department of Human Genetics & Molecular Biology**, University of Health Sciences, Lahore.

Item no.	Item	Concerned bidder	Grievance	Response to grievance
02	Thermal cycler	Analytical Measuring System (AMS)	Review the disqualification for the said item	The bid was disqualified originally because the bidder did not provide relevant information regarding sample ramp rate and insulated metal blocks in their technical bid. However, after reviewing the updated relevant information provided by bidder, it may be considered qualified for the said item.
02	Thermal cycler	Pak Glorious Enterprise (PGE)	Review the disqualification for the said item	The bid was disqualified originally because the bidder did not provide relevant information regarding sample ramp rate and insulated metal blocks in their technical bid. However, after reviewing the updated relevant information provided by bidder, it may be considered qualified for the said item.
25	Biological safety cabinet	Pak Glorious Enterprise (PGE)	Status "not quoted" given in technical evaluation report for the said item	The bid was missed during the preparation of technical evaluation report. However, after review, the bid has been included and is to be considered <b>qualified</b> for the said item.
16	CO <sub>2</sub> incubator	Muslim Trading Agencies (MTR)	Review the disqualification for the said item	The bid was disqualified originally because the bidder failed to provide clearly the relevant information regarding fanless chamber and need of HEPA filters in their technical bid. However, after reviewing the updated relevant information provided by bidder, it may be considered <b>qualified</b> for the said item.
19	Incubation oven	ScienTec	Status "not quoted" given in technical evaluation report for the said item	The bid was missed during the preparation of technical evaluation report. However, after review, the bid has been included and is to be considered <b>qualified</b> for the said item.

Best regards,

Prof. Dr. Shagufta Khaliq

Head.

Dept. of Human Genetics & Molecular Biology, University of Health Sciences, Lahore.

The Incharge Purchase Cell
University of Health Sciences, Lahore

Subject: RE; Submission of grievance against technical evaluation reports

In response to letter from project coordinato (SARF-KSK project) dated 23-07-19, the quotation and specifications mentioned on manufectures's website of the following instruments were evaluated again and it is confirmed that evaluation report is correct as under:

Item	Griervance	UHS requirement (BOQ)	Company name	Specifications Quotation	Comments/ decision
Item ser No. 8, magnetic sterrer with hotplate	The status is not mentioned	*****	Scientec		Qualified
Item Ser No. 22 Co2 incubator	Comparative show not quoted		Muslim trading agency	Quted and found as per BOQ	Qualified
Item Ser No. 22 Co2 incubator	Comparative show specs not mentioned		Muslim trading agency	Some features quted and found as per BOQ	Qualified



Item ser No. 34	Not as per	Inner chamber stainless steel,	Muslim	Not mentioned (both in qutaion and	Disqualified
Incubator	specs	factory set on +37°C	trading	website)	
		Temperature sensor PT100 class	agency		
		A, Temperature setting increments			
		0.1 °C, Timer setting timer 0-999			
		minutes / 0-99.9 hours or			
		continuous, Microprocessor			
		controlled standard			
		microprocessor control with large			
		LED display indications for			
		temperature, adjustable ventilation			
		flap located at the back of the		- "	
		machine			

Item ser no. 34, mentioned instruments does not meet the basic specifications mentioned in BOQ. Therefore, the submitted grievance by Muslim trading agency is inappropriate (print of their website brochure and specs mentioned in quotation are attached as refrence). Rest of the grievances submitted by Muslim trading agency and Scientec has been addressed.

Prof. Dr. Nadeeem Afzal

Head Department of Immunology

University of Health Sciences Lahore.

Specs in BOQ	Specs in Quotation
Heating system direct heat type (dry wall) natural convection Inner chamber stainless steel interior Inner door interior glass door Door mounted control panel Standard Temperature range +5°C above ambient to +70°C factory set on +37°C Temperature sensor PT100 class A Temperature setting increments 0.1 °C Timer setting timer 0-999 minutes / 0-99.9 hours or continuous Microprocessor controlled standard microprocessor control with large LED display indications for temperature, heating and timer Maximum load per shelve 10 kg Temperature deviation +0.6 °C as a function of space Temperature deviation +/- 0.2 °C as a function of time Heating up time to +37°C 42 minutes External dimensions W x D x H mm 610 x 670 x 680 Internal dimensions W x D x H mm 400 x 330 x 410 Shelves 2 chrome-plated grids standard (max 4) W x D mm 595 x 390 220-240V, 50/60 Hz Average power consumption 0.15 kWh Net weight 47 kg Shipping weight and dimensions 61 kg	Incubator Model: BD-s-56 Make: Binder Germany  • • Temperature range: ambient temperature +5 °C to 70 °C • APT.line™ preheating chamber technology • Natural convection • Adjustable exhaust air flap • Controller with timer function • Inner door made of tempered safety glass • 1 chrome-plated rack, incl. shelf supports • Class 3.1 integrated independent temperature safety device (DIN 12880) with visual alarm
Net weight 47 kg	

Website technical sheet

BDS056-230V

BDS056UL-120V

## Designation

Option model	Standard	Standard
Order number	9090-0016	9090-0017
Performance Data Temperature		
Recovery time after 30 seconds door open at 37 °C [min]	15	15
Temperature fluctuation at 37 °C [± K]	0.1	0.1

emperature range 5 °C above ambient temperature to [°C]	70	70	
emperature variation at 37 °C [± K]	0.7	0.7	
Electrical data			
Nominal power [kW]	0.3	0.3	
Phase (Nominal voltage)	1~	1~	
Power frequency [Hz]	50/60	60	
Rated Voltage [V]	230	120	
Unit fuse [A]	6.3	12.5	
Doors			
Inner doors	1	1	
Unit doors	1	1	
Housing dimensions not incl. fittings and connections			
Depth net [mm]	577	577	
Height net [mm]	625	625	
Width net [mm]	614	614	
Internal Dimensions			
Depth [mm]	350	350	
Height [mm]	440	440	
Width [mm]	400	400	
Measures			
		62	

Load per rack [kg]	15	15
Net weight of the unit (empty) [kg]	35.5	35.5
permitted load [kg]	30	30
Wall clearance back [mm]	160	160
Wall clearance sidewise [mm]	100	100
Environment-specific data		
Energy consumption at 37 °C [Wh/h]	25	25
Fixtures		
Number of shelves (std./max.)	1/3	1/3