


TECHNICAL EVALUATION REPORT (SUMMARY)

1. Name of Procuring Agency: University of Health Sciences Lahore
2. Method of Procurement: Single Stage two Envelopes
3. Title of Procurement: Procurement of Laboratory Chemicals/ Consumable for PhD Research Project & Equipment for Human Genetics Department, University of Health Sciences, Lahore
4. Tender Inquiry No: UHS/Pur/1060/2021
5. Date & Time of Bid Closing: 28-01-2021 at 11:00 AM
6. Date & Time of Bid Opening 28-01-2021 at 11:30 AM
7. No. of bids received 09

EQUIPMENT

Sr. #	Item	Specifications	QT N	AGBL Pakistan	Sure Bio Diagnostics	HEK Enterprises	Pak Engimed	Muslim Medical	New Chemical Centre	WWS	PHILCO	SU Enterprises
1	PCR work station	Biobase/thermo/equalent Stainless steel, Internal width 3.5-4 ft, LED display, UV sterilization, Universal caster with leveling feet.	01	Not Qualified Not quoted according to required specifications (working area smaller than required & universal caster quoted as optional)	Not quoted	Qualified	Qualified	Not quoted	Qualified	Not Qualified Not quoted according to required specifications (working area smaller than required)	Not quoted	Qualified
2	Vortex mixer	Genie/VWR/Biobase/equalent Variable speed control, Tube and plate attachments, continuous and touch operations, heavy casted feet.	01	Qualified	Not Qualified Not quoted according to required specifications (plate attachment not available)	Not Qualified Not quoted according to required specifications (plate attachment not available)	Not quoted	Not quoted	Qualified	Qualified	Not quoted	Qualified

3	Pipette set	Gilson/eppendorf/thermo/equivalent 0.1-2.5ul, 0.5-10ul, 10-100ul and 100-1000ul.	01 set of 04 pipettes	Qualified	Qualified	Qualified	Not quoted	Not quoted	Not quoted	Qualified	Qualified	Qualified	
4	Microplate centrifuge	Thermo/Silogics/equivalent Quick spin, Swing out rotor, Compatible with all types (skirted, semi or non) of plates.	01	Not quoted	Qualified	Not quoted	Not quoted	Not quoted	Not quoted	Not Qualified Not quoted according to required specifications (Not swing out rotor)	Qualified	Not Qualified Not quoted according to required specifications (Not swing out rotor)	Qualified
5	Ultra low freezer	Upright, Temperature -80C, Capacity 300-400L, 230V, Climate class N-T.	01	Not quoted	Qualified	Not Qualified Not quoted according to required specifications (Climate class not N-T)	Not quoted	Qualified	Qualified	Qualified	Qualified	Not quoted	Not Qualified Not quoted according to required specifications (Climate class not N-T)



Prof. Dr. Shagufta Khaliq
HOD, Human Genetics