

MANDATORY ELIGIBILITY CRITERIA FOR TECHNICAL EVALUATION

COMPARATIVE STATEMENT :

1. Name of Procuring Agency: University of Health Sciences Lahore
2. Method of Procurement: Single Stage Two Envelope Bidding Procedure under PPRA Rules 2014 (amended)
3. Title of Procurement: Framework contract for procurement of printing of answer books, transcripts, envelopes award lists and misc. Printing of examination department, University of Health Sciences, Lahore for the FINANCIAL YEAR 2025-26 Tender No: UHS/PUR/FW/25/4077
4. Date & Time of Bid Closing: 05-12-2025 - 03:00 PM
5. Date & Time of Bid Opening: 05-12-2025 - 03:30 PM
6. No. of bids received: 04

Sr #	Required Item(s)	Specifications	Unit	Total Quantity	Bidders Name			
					1	2	3	4
					M/s Global Clinical Cura (Pvt) Ltd	M/s Popular International Pvt Ltd.	M/s Roche Pakistan Limited	M/s Reaction Scientific

LOT – B - Framework Contract for the Procurement of Special Chemistry Analyzer with Reagents

1	Special Chemistry Analyzer	<p>IMMUNOASAY ANALYZER</p> <p>TECHNICAL SPECIFICATIONS Detailed Requirement:</p> <ul style="list-style-type: none"> •Fully automatic Random Access Analyzer for routine and specialized immunoassays •Chemiluminescence based System •Qualitative and quantitative analysis assays •Onboard reagent Refrigeration •Automatic Rerun Facility for out of range results •Bar Code Reader for Reagents and Samples •The reagent vials with caps close automatically, no reagent evaporation •Automatic sample clot detection and rejection in each sample to ensure every sample is clear •Automated operation, no manual loading of test units / test cartridges etc. •RS 232 or USB Interface •220V 50 Hz, AC •Quoted equipment should be CE/EU marked and / or US/FDA/IVD approved having external proficiency testing survey enrollment. •Closed System Chemistry Analyzer <p>User Adjustable Settings:</p> <ul style="list-style-type: none"> •Sample type: Serum , plasma or urine •Sample volume: 150 to 250 ul test dependent •Through put min 80 tests/hr or more •Sample Capacity should be at least 50 samples with continuous loading platform •Up to 24 Reagents on-board at a time <p>Displayed Parameters:</p> <ul style="list-style-type: none"> •Automatic assays of analytes related to Cancer, •Infectious diseases, Cardiovascular disease, •Thyroid disease, Fertility/Hormones, Polio & Vitamins •Hepatitis markers <p>Accessories:</p> <ul style="list-style-type: none"> Complete with standard and operation accessories; •Built-in or External Laser Printer •Imported Compatible Sine wave UPS for back up of up to 30 minutes (Emerson, Liebert, Chloride, MGE, APC or Equivalent) •Operating Manual with a Soft Copy •Service Manual with a Soft Copy 	No.	1	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Fully Automated Chemiluminescence Analyzer Model: HUMACLIA 150 With Computer System & UPS	Not Qualified / Non Responsive due to Rate are quoted in Technical Bid.	E411	Qualified / Responsive	CL-900i	Not Qualified / Non responsive due to submission of incomplete Technical Bid i.e. did not mentioned required tests in technical specifications
---	----------------------------	--	-----	---	---	------------------------	--	---	------	------------------------	---------	--

List of Reagents required for Special Chemistry Analyzer

Sr#	Parameters	Unit	Qty							
2	AFP	Set	300	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Human Germany	Not Qualified / Non Responsive due to Rate are quoted in Technical Bid.	AFP	Qualified / Responsive	AFP
3	Beta HCG	Set	300	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Beta HCG	Qualified / Responsive	Beta HCG
4	CA125	Set	300	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		CA125	Qualified / Responsive	CA125
5	Estradiol	Set	300	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Estradiol	Qualified / Responsive	Estradiol
6	Ferritin	Set	300	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Human Germany	Ferritin	Qualified / Responsive	Ferritin	
7	Free T3	Set	300	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Human Germany	Free T3	Qualified / Responsive	Free T3	
8	Free T4	Set	300	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	Human Germany	Free T4	Qualified / Responsive	Free T4	

Not Qualified / Non responsive due to submission of incomplete Technical Bid i.e. did not mentioned required tests in technical specifications

Handwritten signature and checkmark

Handwritten signature and checkmark

Sr #	Required Item(s)	Specifications	Unit	Total Quantity	Bidders Name							
					1		2		3		4	
					M/s Global Clinical Cura (Pvt) Ltd		M/s Popular International Pvt Ltd.		M/s Roche Pakistan Limited		M/s Reaction Scientific	
9	FSH		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany	Not Qualified / Non Responsive due to Rate are quoted in Technical Bid.	FSH	Qualified / Responsive	FSH	Not Qualified / Non responsive due to submission of incomplete Technical Bid i.e. did not mentioned required tests in technical specifications
10	LH		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		LH	Qualified / Responsive	LH	
11	Prolactin		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Prolactin	Qualified / Responsive	Prolactin	
12	PSA		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		PSA	Qualified / Responsive	PSA	
13	Testosterone		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Testosterone	Qualified / Responsive	Testosterone	
14	Total T3		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Total T3	Qualified / Responsive	Total T3	
15	Total T4		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Total T4	Qualified / Responsive	Total T4	
16	TSH		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		TSH	Qualified / Responsive	TSH	
17	Vitamin D		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Vitamin D	Qualified / Responsive	Vitamin D	
18	Controls		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Controls	Qualified / Responsive	Controls	
19	Calibrators		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Calibrators	Qualified / Responsive	Calibrators	
20	Buffers		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Buffers	Qualified / Responsive	Buffers	
21	Plastic Ware		Set	300	Shenzhen New Industries, Biodemical Engineering Co, Ltd China	Qualified / Responsive	Human Germany		Plastic Ware	Qualified / Responsive	Plastic Ware	


Qualified / Responsive	Not-Qualified / Non-Responsive
M/s Global Clinical Cura (Pvt) Ltd	M/s Popular International Pvt Ltd.
M/s Roche Pakistan Limited	M/s Reaction Scientific

TECHNICAL COMMITTEE:

Technical committee met on 09.02.26 to consider the Technical Evaluation Report following single stage two envelope bidding procedure submitted by End User. All bid(s) were technically evaluated according to BOQs and eligibility criteria. Technical Committee recommends to process further as per PPRA Rules 2014 (Amended).


End User
 Co-Opted (Member)


Technical Officer(SHC&ME)
 Member


Dr. Muhammad Imran
 Member


Dr. Saba Khaliq
 Co-Convener


Dr. Uruj Zehra
 Convener



UNIVERSITY OF HEALTH SCIENCES LAHORE

Minutes of the Meeting of Grievance Redressal Committee

A meeting of the Grievance Redressal Committee was held on **Friday, 13th February 2026**, at **02:30 PM** in the Senate Hall, University of Health Sciences, Lahore.

The following participated in the meeting.

- | | |
|--|-------------|
| 1. Dr. Asim Pervaiz | (Convener) |
| Assistant Professor, Department of Human Genetics & Molecular Biology, UHS | |
| 2. Mr. Muhammad Atif | (Member) |
| Director Media & Publications, UHS | |
| 3. Dr. Hummad Hussain | (Member) |
| Additional Registrar (Statutory Meetings), UHS | |
| 4. Mr. Hassan Abidi | (Member) |
| Deputy Registrar, Legal, UHS | |
| 5. Dr. Saima Suleman | (Co-opted) |
| In charge, Biochemistry Department, UHS | |
| 6. Dr. Shabbir Hussain | (Co-opted) |
| Assistant Professor, Institute of Allied Health Sciences, UHS | |
| 7. Mr. Sufyan Ali | Co-opted |
| Secretary, Technical Committee, UHS | |
| 8. Mr. Muhammad Shahzad | (Co-opted) |
| Representative of the Purchase Department) | |
| 9. Mr. Waqas Latif | (Secretary) |
| Data Analyst, QEC, UHS | |

Agenda:

1. Framework contract for the procurement of routine & special Chemistry Analyzers with reagents for Chemical Pathology, University of Health Sciences, Lahore for the Financial Year 2025-26

The Committee was apprised that five (05) grievances were received through the E-PADS Portal regarding the technical evaluation of the above-mentioned tender. The tender was published in accordance with the provisions of the Punjab Procurement Regulatory Authority (PPRA) Rules, 2014 (as amended). The Technical Evaluation Report had been uploaded on the PPRA Portal after scrutiny in light of the mandatory eligibility and technical criteria prescribed in the bidding documents.

The grievances were taken up one by one and deliberated in detail as under:

1. Grievance by M/s Innotech International

The Committee was informed that M/s Innotech International submitted a grievance against the decision declaring their firm non-responsive on the grounds of insufficient experience.

The petitioner contended that certain clauses in the mandatory eligibility criteria regarding experience were conflicting and that the experience requirement applied by the Technical Committee was excessive.

The Committee examined the bidding documents, particularly the Mandatory Eligibility Criteria, which clearly required 10–15 years of relevant experience in the supply of similar nature equipment. The experience documents submitted by the firm were reviewed and re-verified. Upon scrutiny, it was observed that the firm did not meet the minimum required experience as stipulated in the bidding document.

Decision:

The Committee unanimously resolved that the Technical Evaluation Committee had rightly declared the firm non-responsive in accordance with the bidding document. The grievance was, therefore, dismissed.

2. Grievance by M/s Reaction Scientific Pvt. Ltd. (Regarding Non-Quotation of Item No. 27 – Lot-A)

The Committee considered the grievance submitted by M/s Reaction Scientific Pvt. Ltd., wherein the firm requested reconsideration of its bid for Lot-A. The petitioner submitted that Item No. 27 (plastic ware) was not quoted separately as it constituted built-in consumables supplied with the analyzer and, therefore, should not render the bid incomplete.

The Committee reviewed the technical bid submitted by the firm along with the bidding document guidelines. It was clearly stipulated in the bidding document that the bidder must quote and be responsive/qualified in all items included in the LOT/Category to qualify. Upon scrutiny, it was observed that Item No. 27 was not quoted in the technical documents submitted by the firm. The Committee noted that as per PPRA Rules and the explicit condition mentioned in the bidding documents, non-quotation of any listed item in the LOT renders the bid non-responsive.

Decision:

The Committee unanimously resolved that the non-quotation of Item No. 27 constituted non-compliance with the bidding document requirements. The grievance was dismissed.

3. Grievance by M/s Reaction Scientific Pvt. Ltd. (Regarding Kit Size – 300 Tests)

The Committee further considered another grievance submitted by M/s Reaction Scientific Pvt. Ltd., concerning the qualification status related to kit/test quantity. The petitioner submitted that they had quoted the required test kits; however, an issue arose regarding the pack size.

Handwritten signature in blue ink.

Handwritten mark in blue ink.

The Committee examined the technical specifications and BOQ, wherein the required test was specified as 300 tests. Upon scrutiny of the submitted technical bid, it was observed that the firm had quoted a kit size of 100 tests per kit and had not quoted the required 300-test as specified in the bidding document. The bidding guidelines clearly required that the bidder must quote and be responsive/qualified in all items included in the LOT/Category to qualify.

Decision:

Since the firm did not quote the required 300-test configuration as per the technical specifications, the Committee held that the bid was non-responsive. The grievance was dismissed.

4. Grievance by M/s Popular International Pvt. Ltd. Against M/s Global Clinical Cura (Regarding Country of Origin)

The Committee considered a grievance submitted by M/s Popular International Pvt. Ltd. alleging that the quoted model by M/s Global Clinical Cura did not meet the country of origin/manufacturing requirements and that the authorization documentation required verification.

The Committee reviewed the technical documents, authorization letters, and certifications submitted by M/s Global Clinical Cura. It was noted that while the city of origin was not explicitly mentioned in certain documents, the bidding document primarily required compliance with mandatory certifications, including valid FDA/CE certification, as per PPRA Rules and tender conditions. The Committee thoroughly reviewed and re-verified the submitted certifications and found that M/s Global Clinical Cura possessed valid and verifiable FDA/CE certification in compliance with the tender requirements.

Decision:

The Committee unanimously resolved that M/s Global Clinical Cura met the mandatory certification requirements. The grievance was, therefore, dismissed.

5. Grievance by M/s Popular International Pvt. Ltd. Against M/s Global Clinical Cura (Regarding Experience Requirement)

The Committee further considered another grievance submitted by M/s Popular International Pvt. Ltd. alleging that M/s Global Clinical Cura did not fulfill the required 10–15 years' experience as prescribed in the mandatory eligibility criteria.

The Committee re-verified the experience documents submitted by M/s Global Clinical Cura, including relevant registrations and credentials. The experience details were cross-checked through the FBR portal and supporting documentation. Upon verification, it was found that

21/2/2026

the firm fulfilled the required minimum 10 years' relevant experience as stipulated in the bidding document.

Decision:

The Committee unanimously resolved that the Technical Evaluation Committee had correctly assessed the experience credentials of M/s Global Clinical Cura. The grievance was, therefore, dismissed.

Conclusion:

After detailed deliberation and examination of all relevant documents, the Grievance Redressal Committee unanimously resolved that all five (05) grievances received regarding the Framework Contract for the Procurement of Routine & Special Chemistry Analyzers with Reagents for Chemical Pathology for the Financial Year 2025-26 were devoid of merit. Accordingly, all grievances were dismissed, and the Technical Evaluation Report stands upheld as per PPRA Rules 2014 (as amended) and the provisions of the bidding documents. The meeting concluded with a vote of thanks to and from the Chair.



Dr. Asim Pervaiz
(Convener)
Assistant Professor, Department of
Human Genetics & Molecular
Biology, UHS




Mr. Muhammad Atif
(Member)
Director Media &
Publications, UHS



Dr. Hummad Hussain
(Member)
Additional Registrar (Statutory
Meetings), UHS



Mr. Hassan Abidi
(Member)
Deputy Registrar (Legal), UHS



Mr. Waqas Latif
(Secretary)
Data Analyst, QEC, UHS

Dr. Saima Suleman
(Co-opted)
In charge, Biochemistry
Department, UHS



Dr. Shabbir Hussain
(Co-opted)
Assistant Professor, Institute of Allied
Health Sciences, UHS



Mr. Sufyan Ali
(Co-opted)
Secretary, Technical
Committee, UHS



Mr. Muhammad Shahzad
(Co-opted)
Representative of the
Purchase Department)

MANDATORY ELIGIBILITY CRITERIA FOR TECHNICAL EVALUATION

COMPARATIVE STATEMENT :


1. Name of Procuring Agency: University of Health Sciences Lahore
2. Method of Procurement: Single Stage Two Envelope Bidding Procedure under PPRA Rules 2014 (amended)
3. Title of Procurement: Framework Contract for the Procurement of Routine & Special Chemistry Analyzers with Reagents for Chemical Pathology, University of Health Sciences, Lahore for The FINANCIAL YEAR 2025-26 Tender No: UHS/PUR/FW/25/147
4. Date & Time of Bid Closing: 03-12-2025 - 03:00 PM
5. Date & Time of Bid Opening: 03-12-2025 - 03:30 PM
6. No. of bids received: 07


Sr #	Items with specifications	Bidder Name						
		1	2	3	4	5	6	7
		M/s Alfa Scientific Store	M/s Global Clinical Cura (Pvt) Ltd	M/s Innotech International	M/s NexTek Health Care	M/s Popular International Pvt Ltd.	M/s Reaction Scientific	M/s Roche Pakistan Limited
1	Copy of NTN /GST	✓	✓	✓	✓	✓	✓	✓
2	Copy of Professional Tax	✓	✓	✓	✓	✓	✓	✓
3	Registered with Tax Department(s): Declaration issued by Tax Department, Registration with tax department / Tax Returns/ Tax paid / Active tax payer status in FBR record	✓	✓	✓	✓	✓	✓	✓
4	Copy of Bank Guarantees / CDR / Pay Order per Bid Security	✓	✓	✓	✓	✓	✓	✓
5	Bank Account Maintenance: Fresh Certificate issued by bank that the account is being maintained since when and the title of account is what	✓	✓	✓	✓	✓	✓	✓
6	Last 01 year bank account statement	✓	✓	✓	✓	✓	✓	✓
7	Undertaking on Stamp Paper for non-blacklisting by the procuring agency	✓	✓	✓	✓	✓	✓	✓
8	The bidder must have 10 to 15 years' experience in relevant business	✓	✓	19-11-2019 (06Y)	21-10-2017 (08Y-01M)	✓	✓	✓
9	A copy of a valid authorization certificate from the primary / principle manufacturer is required. If the evidence of manufacturing is provided by the manufacturer, it should be included as well. (if applicable)	✓	✓	Sub Dealer	✓	✓	✓	✓
10	Upload copy of bid document of UHS with each page signed and stamped	✓	✓	✓	✓	✓	✓	✓
11	Portfolio of Firm (supporting documents should be attached)	✓	✓	✓	✓	✓	✓	✓
12	The Bidder shall submit at least five (05) purchase/work orders in a Govt / Public Sector Institute	✓	✓	✓	✓	✓	✓	✓
13	Quoted equipment should be CE/EU marked and / or US/FDA/IVD approved having external proficiency testing survey enrollment.	X	✓	✓	✓	✓	✓	✓

It is hereby confirmed that three bids were evaluated against the laid specifications and found four bids qualified/responsive in accordance with the submitted documents. Comparative statement (summary/detail) is being submitted to Technical Committee, UHS for further necessary action.


TECHNICAL COMMITTEE:


Technical committee met on 22-01-26 to consider the Technical Evaluation Report following single stage two envelopes bidding procedure submitted by End User. Four bid(s) were technically evaluated according to BOQs and eligibility criteria. Technical Committee recommends to process further as per PPRA Rules 2014 (Amended).


End User
Co-Opted (Member)


Technical Officer(SHC&ME)
Member


Dr. Muhammad Imran
Member


Dr. Saba Khaliq
Co-Convener


Dr. Uruj Zehra
Convener

MANDATORY ELIGIBILITY CRITERIA FOR TECHNICAL EVALUATION

COMPLIANCE STATEMENT :

1. Name of Procuring Agency: University of Health Sciences Lahore
2. Method of Procurement: Single Stage Two Envelope Bidding Procedure under PPRA Rules 2014 (amended)
3. Title of Procurement: Framework contract for procurement of printing of answer books, transcripts, envelopes award lists and misc. Printing of examination department, University of Health Sciences, Lahore for the FINANCIAL YEAR 2025-26 Tender No: UHS/PUR/FW/25/4077
4. Date & Time of Bid Closing: 05-12-2025 - 03:00 PM
5. Date & Time of Bid Opening: 05-12-2025 - 03:30 PM
6. No. of bids received: 07


Sr #	Required Item(s)	Specifications	Unit	Total Quantity	Bidders Name												
					1	2	3	4	5	6	7						
<p>LOT - A - Framework Contract for the Procurement of Routine Chemistry Analyzer with Reagents</p>																	
1	Routine Chemistry Analyzer	<p>LOW END FULLY AUTOMATIC RANDOM ACCESS CHEMISTRY ANALYZER</p> <p>TECHNICAL SPECIFICATIONS Detailed Requirements: •Microprocessor controlled fully automated chemistry analyzer. •User programmable with built in bar code reader •Sample type: Serum, Plasma, Body Fluids etc. •Capability to re-run with automatic on board sample dilution. •Automatic Sample Quality Analysis for Lipemia, Hemolysis, Icterus. •CD/USB Storage System •RS 232 interface for on line computer •220V, 50 Hz operated •Quoted equipment should be CE/EU marked and / or US/FDA/IVD approved having external proficiency testing survey enrollment. •Closed System Chemistry Analyzer.</p> <p>User Adjustable Settings: • Minimum 30 tests on panel including ISE •Minimum 35-45 sample/QC positions and continuous loading of samples •130 tests/hr or above for a range of chemistries (Other than ISE i.e. Na+, K+, Cl-) •Spectral Range 320-690 nm •Diffraction grating wavelength 340, 405, 492, 505, 546, 578, 630 nm or more / Filter. Wavelength 340, 405, 492, 505, 546, 578, 630 nm, one free position •Sample volume 2-50 µl</p> <p>Displayed Parameters: • LCD Display Monitor •On board reagent refrigeration of 2-8 Degree Centigrade / On board reagent refrigeration compatible with the kits. •Temperature control for assays at 37C •Self calibrating against known standards + Storage of QC results •Independent stat capability, facility of reflex testing •Automatic calibration of curves and results •Automatic flagging of results outside user defined limits •Data entry by keyboard, bar code reader & LIS •Bubble and sample level detector/Liquid Level Sensing on both Reagent and Sample Probe Minimum data storage 10,000 tests</p> <p>Accessories: •With Built-in Thermal Printer or External Laser Printer •Compatible Imported Online Sine wave UPS with Battery backup for 30 minutes (Emerson, Liebert, Chloride, MGE, APC or Equivalent) •Compatible RO System (If required) to operate the Chemistry Analyzer</p>	No.	1	<p>OFFER : CM 2601 FULLY AUTO CHEMISTRY ANALYZER BY WIENER LAB GROUP (Switzerland / Argentina) 01 Analyzer with UPS Backup</p> <p>CE, ISO, FSC, DRAP</p>	Not Qualified / Non Responsive due to non submission of mandatory documents. i.e. External Quality Control Survey report no submitted	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	<p>Special Chemistry Analyzer Throughput without ISE module 200 test/h and Throughput with ISE module 300 test/h Calorimetry, turbidimetry, UV Methods, ISE method 88 (44 bottles of 20 mL, 60 mL or sample tube + 44 bottles of 20 mL or sample tubes) Yes Diameter from 12 mm to 16 mm (height up to 100 mm) 13.5 mm diameter Ceramic piston pump with low maintenance From 1:1 to 1:200 Inside and outside Yes 20 mL, 60 mL</p>	Not qualified / non responsive due to less experience.	Germany, Respons 240c	Not qualified / non responsive due to less experience.	<p>Fully Automatic Chemistry Analyzer Model: Pictus 500 with ISE CE/FDA Approved</p>	Qualified / Responsive	Not Qualified / Non responsive due to submission of incomplete Technical Bid i.e. not quote in all items	cobas c 311 analyzer	Qualified / Responsive

List of Reagents required for Routine Chemistry Analyzer

Sr#	Parameters	Unit	QTY	Vendors Specification	Not Qualified / Non Responsive due to non submission of mandatory documents. i.e. External Quality Control Survey report no submitted	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Not qualified / non responsive due to less experience.	Diatron /Hungry	Qualified / Responsive	Not Qualified / Non responsive due to submission of incomplete Technical Bid i.e. not quote in all items	Albumin	Qualified / Responsive
2	Albumin	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Not Qualified / Non Responsive due to non submission of mandatory documents. i.e. External Quality Control Survey report no submitted	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Not qualified / non responsive due to less experience.	Diatron /Hungry	Qualified / Responsive	Not Qualified / Non responsive due to submission of incomplete Technical Bid i.e. not quote in all items	Albumin	Qualified / Responsive
3	ALP	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		ALP	Qualified / Responsive
4	ALT	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		ALT	Qualified / Responsive
5	AST	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		AST	Qualified / Responsive
6	Total Bilirubin	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Total Bilirubin	Qualified / Responsive
7	Direct Bilirubin	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Direct Bilirubin	Qualified / Responsive
8	Amylase	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Amylase	Qualified / Responsive
9	Lipase	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Lipase	Qualified / Responsive
10	Creatinine	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Creatinine	Qualified / Responsive
11	Urea	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Urea	Qualified / Responsive
12	Uric Acid	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Uric Acid	Qualified / Responsive
13	Glucose	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Glucose	Qualified / Responsive
14	Total Cholesterol	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP		Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c		Diatron /Hungry	Qualified / Responsive		Total Cholesterol	Qualified / Responsive
15	HDL Cholesterol	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	HDL Cholesterol	Qualified / Responsive			
16	LDL Cholesterol	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	LDL Cholesterol	Qualified / Responsive			
17	Triglycerides	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	Triglycerides	Qualified / Responsive			
18	Total Protein	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	Total Protein	Qualified / Responsive			
19	CKMB	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	CKMB	Qualified / Responsive			
20	CPK	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	CPK	Qualified / Responsive			
21	hsCRP	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	hsCRP	Qualified / Responsive			
22	LDH	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	LDH	Qualified / Responsive			
23	HbA1C	Parameters	2000	(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	HbA1C	Qualified / Responsive			
24	Calibrators	Calibrators, Daily all applicable controls for (6 Months), buffers, and plastic-ware		(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	Calibrators	Qualified / Responsive			
25	Controls	Controls		(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	Controls	Qualified / Responsive			
26	Buffers	Buffers		(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	Buffers	Qualified / Responsive			
27	Plastic ware	Plastic ware		(Switzerland / Argentina) FDA, CE, ISO, FSC, DRAP	Shenzhen New Industries, Biomedical Engineering Co, Ltd China	Qualified / Responsive	BioSystems/Spain	Germany, Respons 240c	Diatron /Hungry	Qualified / Responsive	Plastic ware	Qualified / Responsive			

Qualified / Responsive	Not-Qualified / Non-Responsive
M/s Global Clinical Cura (Pvt) Ltd	M/s Reaction Scientific
M/s Popular International Pvt Ltd.	M/s NexTek Health Care
M/s Roche Pakistan Limited	M/s Innotech International
	M/s Alfa Scientific Store

TECHNICAL COMMITTEE: Technical committee met on 22-12-26 to consider the Technical Evaluation Report following single stage two envelope bidding procedure submitted by End User. All bid(s) were technically evaluated according to BOQs and eligibility criteria. Technical Committee recommends to process further as per PPRA Rules 2014 (Amended).



 End User Co-Opted (Member) Technical Officer(SHC&ME) Member Dr. Umera Saleem Member Dr. Saba Khalid Co-Convenor Dr. Uruj Zehra Convenor